

## FDA Recall – Cardinal Health

### Purpose of this communication:

- We are writing to inform you that Cardinal Health has issued a voluntary recall for select lots of Webcol™ Large Alcohol Prep Pads (70% isopropyl alcohol) to the consumer level. The product is being recalled due to microbial contamination identified as *Paenibacillus phoenicis*.

### What do I need to do?

- Please review the following recall notice: [https://www.fda.gov/safety/recalls-market-withdrawals-safety-alerts/cardinal-health-issues-voluntary-nationwide-recall-webcoltm-large-alcohol-prep-pad?utm\\_medium=email&utm\\_source=govdelivery](https://www.fda.gov/safety/recalls-market-withdrawals-safety-alerts/cardinal-health-issues-voluntary-nationwide-recall-webcoltm-large-alcohol-prep-pad?utm_medium=email&utm_source=govdelivery)
- Notify impacted patients and facilitate the replacement, according to the guidelines issued by the manufacturer in the FDA notification.
- Confirm if you have any CareCentrix patients affected by this recall and notify the CareCentrix Quality Department by faxing information to: 919-792-6718 Attn: Quality Department.

Thank you in advance for your cooperation and continued partnership.