

## FDA Recall – Philips Respironics DreamStation Devices

### Purpose of this communication:

We are writing to inform you that Philips Respironics is recalling certain DreamStation Auto CPAP and Auto BiPAP devices due to a programming error introduced during rework by a supplier, which may result in incorrect therapy modes (e.g., BiPAP configured as CPAP), limited pressure, unavailable features, or improper sensor/humidifier function. Use of the affected devices may cause serious health consequences including hypoventilation, disrupted sleep, skin or airway burns, and ineffective treatment of sleep apnea, which could worsen comorbidities. There have been three reported injuries and no reports of death.

### What do I need to do?

- Please review the following recall notice [here](#).
- Notify impacted patients and facilitate the replacement, according to the guidelines issued by the manufacturer in the FDA notification.
- Confirm if you have any CareCentrix patients affected by this recall and notify the CareCentrix Quality Department by faxing information to: 919-792-6718 Attn: Quality Department.

Thank you in advance for your cooperation and continued partnership.